



宁波柯力传感科技股份有限公司  
KELI Sensing Technology (Ningbo) Co., Ltd

Certificate number: TR-LCKL-2412005

Page 1 of 2

### CERTIFICATE OF CONFORMITY

#### Manufacturer of the certified and Issuing authority

Name : Keli Electric Manufacturing (Ningbo) Co.,Ltd.

Address : No. 199 Changxing Road,  
Jiangbei District, Ningbo City  
P.R. of China

#### Distributor of the certified type

Name : Digital Scale & Engineering Co.,Ltd

Address : 8 Charoen Nakhon 35 Alley,  
Bang Lamphu Lang, Khlong San,  
Bangkok 10600

#### Identification of certified type

Type : ZSFY-A

Fraction :  $P_i = 0.7$

Temperature range  $-10^{\circ}\text{C} / 40^{\circ}\text{C}$

Maximum capacity ( $E_{\max}$ )	10 000 kg up to and including 30 000 kg
Minimum dead load ( $E_{\min}$ )	$0\% * E_{\max}$
Accuracy Class	C
Rated Output (RO)	$2.0 \pm 0.002 \text{ mV/V}$
Maximum number of load cell intervals (n)	4000
Ratio of minimum LC Verification interval $Y = E_{\max} / V_{\min}$	10 000
Ratio of minimum dead load output return $Z = E_{\max} / (2 * DR)$	3000
Input impedance	$700 \pm 20 \Omega$

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report ) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R60**

Edition 2000 ( E )

for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The Issuing Authority



沈岳

DATE: 2023-12-14