



宁波柯力传感科技股份有限公司  
KELI Sensing Technology (Ningbo) Co., Ltd

Certificate number: TR-LCKL-2412002

Page 1 of 2

**CERTIFICATE OF CONFORMITY**

**Manufacturer of the certified and Issuing authority**

Name : Keli Electric Manufacturing (Ningbo) Co.,Ltd.  
Address : No. 199 Changxing Road,  
Jiangbei District, Ningbo City  
P.R. of China

**Distributor of the certified type**

Name : Digital Scale & Engineering Co.,Ltd  
Address : 8 Charoen Nakhon 35 Alley,  
Bang Lamphu Lang, Khlong San,  
Bangkok 10600

**Identification of certified type**

Type : EDS-D  
Fraction :  $P_i = 0.8$   
Temperature range  $-30^{\circ}\text{C} / 70^{\circ}\text{C}$

Maximum capacity ( $E_{\max}$ )	10 000 kg up to and including 40 000 kg
Minimum dead load ( $E_{\min}$ )	0kg
Accuracy Class	C
Maximum number of load cell intervals (n)	5000
Ratio of minimum LC Verification interval $Y = E_{\max} / V_{\min}$	8000
Ratio of minimum dead load output return $Z = E_{\max} / (2 * DR)$	3000

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report ) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R60**

Edition 2000 ( E )  
for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The Issuing Authority

DATE: 2023-12-14



沈岳